



UNITED STATES PATENT AND TRADEMARK OFFICE

UNITED STATES DEPARTMENT OF COMMERCE
United States Patent and Trademark Office
Address: COMMISSIONER FOR PATENTS
P.O. Box 1450
Alexandria, Virginia 22313-1450
www.uspto.gov

APPLICATION NO.	FILING DATE	FIRST NAMED INVENTOR	ATTORNEY DOCKET NO.	CONFIRMATION NO.
09/870,583	06/01/2001	Masashi Ishida	NPR-073	9221

20374 7590 02/19/2004

KUBOVCIK & KUBOVCIK
SUITE 710
900 17TH STREET NW
WASHINGTON, DC 20006

EXAMINER

BAXTER, JESSICA R

ART UNIT	PAPER NUMBER
----------	--------------

3731

DATE MAILED: 02/19/2004

Please find below and/or attached an Office communication concerning this application or proceeding.

Office Action Summary

Applicati n No.

09/870,583

Applicant(s)

ISHIDA ET AL.

Examiner

Jessica R Baxter

Art Unit

3731

-- The MAILING DATE of this communication appears on the cover sheet with the correspondence address --

Period for Reply

A SHORTENED STATUTORY PERIOD FOR REPLY IS SET TO EXPIRE 3 MONTH(S) FROM THE MAILING DATE OF THIS COMMUNICATION.

- Extensions of time may be available under the provisions of 37 CFR 1.136(a). In no event, however, may a reply be timely filed after SIX (6) MONTHS from the mailing date of this communication.
- If the period for reply specified above is less than thirty (30) days, a reply within the statutory minimum of thirty (30) days will be considered timely.
- If NO period for reply is specified above, the maximum statutory period will apply and will expire SIX (6) MONTHS from the mailing date of this communication.
- Failure to reply within the set or extended period for reply will, by statute, cause the application to become ABANDONED (35 U.S.C. § 133). Any reply received by the Office later than three months after the mailing date of this communication, even if timely filed, may reduce any earned patent term adjustment. See 37 CFR 1.704(b).

Status

- 1) ☒ Responsive to communication(s) filed on 01 December 2003.
- 2a) ☐ This action is **FINAL**. 2b) ☒ This action is non-final.
- 3) ☐ Since this application is in condition for allowance except for formal matters, prosecution as to the merits is closed in accordance with the practice under *Ex parte Quayle*, 1935 C.D. 11, 453 O.G. 213.

Disposition of Claims

- 4) ☒ Claim(s) 1 is/are pending in the application.
- 4a) Of the above claim(s) _____ is/are withdrawn from consideration.
- 5) ☐ Claim(s) _____ is/are allowed.
- 6) ☒ Claim(s) 1 is/are rejected.
- 7) ☐ Claim(s) _____ is/are objected to.
- 8) ☐ Claim(s) _____ are subject to restriction and/or election requirement.

Application Papers

- 9) ☐ The specification is objected to by the Examiner.
- 10) ☐ The drawing(s) filed on _____ is/are: a) ☐ accepted or b) ☐ objected to by the Examiner.
Applicant may not request that any objection to the drawing(s) be held in abeyance. See 37 CFR 1.85(a).
Replacement drawing sheet(s) including the correction is required if the drawing(s) is objected to. See 37 CFR 1.121(d).
- 11) ☐ The oath or declaration is objected to by the Examiner. Note the attached Office Action or form PTO-152.

Priority under 35 U.S.C. § 119

- 12) ☐ Acknowledgment is made of a claim for foreign priority under 35 U.S.C. § 119(a)-(d) or (f).
- a) ☐ All b) ☐ Some * c) ☐ None of:
1. ☐ Certified copies of the priority documents have been received.
 2. ☐ Certified copies of the priority documents have been received in Application No. _____.
 3. ☐ Copies of the certified copies of the priority documents have been received in this National Stage application from the International Bureau (PCT Rule 17.2(a)).

* See the attached detailed Office action for a list of the certified copies not received.

Attachment(s)

- | | |
|--|---|
| 1) <input type="checkbox"/> Notice of References Cited (PTO-892) | 4) <input type="checkbox"/> Interview Summary (PTO-413)
Paper No(s)/Mail Date. _____ |
| 2) <input type="checkbox"/> Notice of Draftsperson's Patent Drawing Review (PTO-948) | 5) <input type="checkbox"/> Notice of Informal Patent Application (PTO-152) |
| 3) <input type="checkbox"/> Information Disclosure Statement(s) (PTO-1449 or PTO/SB/08)
Paper No(s)/Mail Date _____ | 6) <input type="checkbox"/> Other: _____ |

DETAILED ACTION

Continued Examination Under 37 CFR 1.114

1. A request for continued examination under 37 CFR 1.114, including the fee set forth in 37 CFR 1.17(e), was filed in this application after final rejection. Since this application is eligible for continued examination under 37 CFR 1.114, and the fee set forth in 37 CFR 1.17(e) has been timely paid, the finality of the previous Office action has been withdrawn pursuant to 37 CFR 1.114. Applicant's submission filed on October 29, 2003 has been entered.

Claim Rejections - 35 USC § 102

2. The following is a quotation of the appropriate paragraphs of 35 U.S.C. 102 that form the basis for the rejections under this section made in this Office action:

A person shall be entitled to a patent unless –

(b) the invention was patented or described in a printed publication in this or a foreign country or in public use or on sale in this country, more than one year prior to the date of application for patent in the United States.

3. Claim 1 is rejected under 35 U.S.C. 102(b) as being anticipated by U.S. Patent No. 5,053,017 to Chamuel.

Chamuel discloses a clamp (20') having a proximal side and a distal side comprising: a flexible curved portion (38) on the proximal side having a tube insertion hole (22'); an upper member (26') and a lower member (30') connected via said curved portion, the upper member and the lower member being detachably engaged at distal ends thereof so that a clearance is provided between the upper member and the lower member when said members are not engaged (FIG. 4); a pressurizing portion (Column 3 lines 1-13) provided on said upper member or lower member; and a space (space between end 38 and surface 34') provided in the upper member or lower member; wherein the upper member is provided

Art Unit: 3731

with a pair of left and right engaging hooks (28') at its distal end, and the lower member is provided with an engaged portion.

Response to Arguments

4. Applicant's arguments filed October 29, 2003 have been fully considered but they are not persuasive.

5. In response to applicant's arguments, the recitation "to be placed and used on the tube of a winged needle and for storing the needle after use, the winged needle including a needle cannula, a hub with a wing, and a tube connected to the hub" has not been given patentable weight because the recitation occurs in the preamble. A preamble is generally not accorded any patentable weight where it merely recites the purpose of a process or the intended use of a structure, and where the body of the claim does not depend on the preamble for completeness but, instead, the process steps or structural limitations are able to stand alone. See *In re Hiraio*, 535 F.2d 67, 190 USPQ 15 (CCPA 1976) and *Kropa v. Robie*, 187 F.2d 150, 152, 88 USPQ 478, 481 (CCPA 1951).

6. Applicant argues that the clamp of Chamuel does not disclose that the upper member and lower member are not detachably engaged. However, Chamuel states in Column 3 lines 3-12, that the compression portion is *slightly* longer than the alignment portion and that at some point that the surface 32' and bottom 27' are only separated by a diameter of needle 24 (Column 3 lines 61-65). The guide means 28' will contact the surface 34' when the needle is pulled back relative to the clamp. Therefore, the rejection over Chamuel is proper.

7. In response to applicant's argument that "through which a hub with a wing is introduced when the clamp is disposed on said tube and said hub with a wing is slid onto the clamp" is not met by Chamuel '017, a recitation of the intended use of the claimed invention must result in a structural

Art Unit: 3731

difference between the claimed invention and the prior art in order to patentably distinguish the claimed invention from the prior art. If the prior art structure is capable of performing the intended use, then it meets the claim. In a claim drawn to a process of making, the intended use must result in a manipulative difference as compared to the prior art. See *In re Casey*, 152 USPQ 235 (CCPA 1967) and *In re Otto*, 136 USPQ 458, 459 (CCPA 1963).

8. In response to applicant's argument that Chamuel does not disclose a portion "for engaging and pressurizing said tube when said members are engaged" and a space "for storing said needle cannula and said hub", a recitation of the intended use of the claimed invention must result in a structural difference between the claimed invention and the prior art in order to patentably distinguish the claimed invention from the prior art. If the prior art structure is capable of performing the intended use, then it meets the claim. In a claim drawn to a process of making, the intended use must result in a manipulative difference as compared to the prior art. See *In re Casey*, 152 USPQ 235 (CCPA 1967) and *In re Otto*, 136 USPQ 458, 459 (CCPA 1963).

Conclusion

Any inquiry concerning this communication or earlier communications from the examiner should be directed to Jessica R Baxter whose telephone number is 703-305-4069. The examiner can normally be reached on M-F 8:30AM - 5:00PM.


If attempts to reach the examiner by telephone are unsuccessful, the examiner's supervisor, Michael Milano can be reached on 703-308-2496. The fax phone number for the organization where this application or proceeding is assigned is 703-872-9306.

Art Unit: 3731

Information regarding the status of an application may be obtained from the Patent Application Information Retrieval (PAIR) system. Status information for published applications may be obtained from either Private PAIR or Public PAIR. Status information for unpublished applications is available through Private PAIR only. For more information about the PAIR system, see <http://pair-direct.uspto.gov>. Should you have questions on access to the Private PAIR system, contact the Electronic Business Center (EBC) at 866-217-9197 (toll-free).

Jessica R Baxter
Examiner
Art Unit 3731


jrb


MICHAEL J. MILANO
SUPERVISORY PATENT EXAMINER
TECHNOLOGY CENTER 3700